Supplement

Title: 63/66/68 Users Supplement Issue: **2**Part Number: 2149032 Issue Date: 2/16
Print Date: Sept. 2004 Page Count: 3

Revision/Date: 3, 6/09

This supplement contains information necessary to ensure the accuracy of the above Manual.



Change #1, 46, 401

On page 11, under the *Distance and Spot Size* section add:

For calibrations, hold the thermometer at a distance of 350 mm (14 in) from a target of at least 150 mm (6 in) in diameter.

On page 22, remove the entire **CE Certification** section.

On page 24, remove the **Certifications** section and add:

Electromagnetic Compatibility (EMC)

IEC61326-1 (Portable EM environment)

CISPR11, Group1, Class A

Group 1 equipment: group 1 has intentionally generated and/or use conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.

Class A equipment is equipment suitable for use in all establishments other than domestic and those directly

2/16

connected to a low voltage power supply network which supplies buildings used for domestic purposes.

Note

It may be difficult to ensure electromagnetic compatibility in other environments, due to conducted and radiated disturbances. Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.

The equipment may not meet the immunity requirements of IEC 61326-1 when test leads and/or test probes are connected.

USA (FCC) - 47 CFR 15 subpart B, this product is considered an exempt device per clause 15.103.

Korea (KCC) - Class A Equipment (Industrial Broadcasting & Communication Equipment)

This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

Notes

63: Between 165 MHz and 880 MHz (\pm 5 %) at 3 V/m, the thermometer may not meet its stated accuracy.

66/68: Between 162 MHz and 792 MHz (±5 %) at 3 V/m, the thermometer may not meet its stated accuracy.

Change #2, 239

On page 24, replace Certifications, with:

Certifications: CE, MC

皖制00000287号

On the inside of the back cover of the Manual, add:

执行标准: Q/ASF 07

制造商:安徽世福仪器有限公司

生产地址:安徽省芜湖市鸠江经济开发区龙

腾路66号

电话:0553-5610888